



Department of Botany,
35, Ballygunge Circular Road
Kolkata – 700 019, W.B., INDIA

Notice

Ref. No. BOT/RMI/GL/05/2025-26

Dated. 09.10.2025

Quotations are invited from registered vendors, for the following item. Service providers are there for requested to submit their quotations, including the current Trade License, service Tax Registration Certificate, Pan No., Copy of Registration certificate with statutory authorities and all other credentials within seven days from the publication of notice, in a sealed envelope addressed to the undersigned. The authority reserves the right to accept/reject any/ all quotations without showing any reasons thereof.

DAPI filter and lens Leica Inverted Fluorescence Microscope

ET-DAPI; AT 350/50; ET 460/50m; 400LP without cube

EXTENDED/ENHANCED TRANSMISSION DAPI FILTER SET WITHOUT MOUNTING CUBE
SPECIFICATION:

EXCITATION FILTER

- a. Central Wavelength: 350 nm
- b. Bandwidth (FWHM) = 50 nm
- c. Angle of Incidence = 0°

DICHROIC MIRROR:

- a. Type BS
- b. Central Wavelength: 400 nm
- c. Angle of Incidence = 45°

EMISSION FILTER

- a. Central Wavelength: 460 nm
- b. Bandwidth (FWHM) = 50 nm
- c. Angle of Incidence = 0°

Yours faithfully,
Head, Department of Botany
University Of Calcutta
35, Ballygunge Circular Road
Kolkata-700019