

**UNIVERSITY COLLEGES OF SCIENCE AND TECHNOLOGY
DEPARTMENT OF CHEMISTRY
UNIVERSITY OF CALCUTTA**

Phone : 2350-9937/1014/1857/1397

Fax : 910-33-351-9755

92, Acharya Prafulla Chandra Road

Kolkata : 7000 09, INDIA



Ref. No.CU-Chem/AP/ING/Equip /2025-26/02

Date: 07.11.2025

QUOTATION NOTICE

Quotations are invited from authorized parties for supplying an Analytical Balance with specification detail given below for the **Inorganic Section, Department of Chemistry**. Suppliers are hereby requested to submit their quotations including the current GST information, Trade License, Service Tax Registration Certificate, PAN No., Copy of Registration Certificate with Statutory Authorities and other credentials within **Ten (10) working days** from the publication of this notice on **Calcutta University Website** in a sealed envelope addressed to **The Head, Dept. of Chemistry, University of Calcutta**. The authority reserves the right to accept/reject any quotation without showing any reason thereof. Mentioning **INORGANIC SECTION** at the top of the **ENVELOPE**.

Item: Analytical Balance
Salient Technical Specifications:

Weighing Capacity	220 gm
Minimum Display	0.1mg
Pan Size	Approx. 91 mm dia.
Dimensions	Approx. W213 × D356 × H338 mm
Weight	Approx. 6.2 kg
Required Power Supply	AC adapter (input 100-240 V AC, 50/60 Hz, output 12 V, 1 A)
I/O Terminals	RS232C (D-sub 9P plug), USB device (Type B)
Calibration Weight	Built-in
Calibration	Internal
Warranty	1 year

Kuntal Pal
07/11/25
Dr. Kuntal Pal
Assistant Professor
Department of Chemistry
University of Calcutta

Sanchita Goswami
7/11/25
Departmental purchase committee
member
Dept. of Chemistry
University of Calcutta
Prof. Sanchita Goswami
Professor
Department of Chemistry
University of Calcutta

Nikhil Guchhait
28/11/25
Head of the department
Dept. of Chemistry
University of Calcutta

Prof. Nikhil Guchhait
Head, Department of Chemistry
University of Calcutta