

Dr. Dipankar Chakraborti
Associate Professor
Head of the Department
Department of Genetics
University of Calcutta



University of Calcutta
35, Ballygunge Circular Road
Kolkata – 700 019
Tel: +91-33-2461 5445

QUOTATION NOTICE

Ref. No.: Genet/BI(97)/LERC20/22-23

Dated: 17.10.2022

Quotations for the following item are invited from authorized parties for Department of Genetics. The cost to be incurred from Departmental BI grant. Suppliers are hereby requested to submit their quotation including the copies of current GST registration certificate, PAN card, registration certificate with statutory authorities and other credentials **within 5 working days** from the publication of this notice in a sealed envelope addressed to the undersigned. **Please provide a tender complete in all aspects, otherwise, it won't be considered.** The authority reserves the right to accept/reject any quotation without showing any reason thereof.

No	Material Description.	Qty
1	Digital INCUBATOR WITH SHAKER Specification: Double walled with inner chamber of stainless steel & outer wall of M.S. Sheet duly Powder coated. Temperature will be controlled by Digital controller with air circulation. Temperature range : 5°C to maximum 60°C with accuracy range of $\pm 2^\circ\text{C}$. Chamber Size: 120 lit (4 cft) : Shaker Speed : 20-180 RPM Shaker Size : 13*13 inch Capacity : 500 ml *5 /100 ml*5 & 1 lit *2 Number of shelves: one (1 Nos. Stainless Steel adjustable trays) Microprocessor based PID controller Temperature range between which will be controlled by means of a Digital Temperature Controller in conjunction with an electronic relay.	01

Yours faithfully,

Dr. Dipankar Chakraborti
Head of the Department
Department of Genetics,
University of Calcutta, Kolkata