

Dr. Dipankar Chakraborti
Head of the Department
Department of Genetics
University of Calcutta



University of Calcutta
35, Ballygunge Circular Road
Kolkata – 700 019
Tel: +91-33-2461 5445

QUOTATION NOTICE

Ref. No.: Genet/BI(97)/LERC10/22-23

Dated: 01.08.2022

Quotations for the following item are invited from authorized parties for Department of Genetics. The cost to be incurred from Departmental BI grant. Suppliers are hereby requested to submit their quotation including the current GST registration certificate, PAN no., copy of registration certificate with statutory authorities and other credentials **within 5 working** days from the publication of this notice in a sealed envelope addressed to the undersigned. **Please provide a tender complete in all aspects, otherwise, it won't be considered.** The authority reserves the right to accept/reject any quotation without showing any reason thereof.

Item	Description	Qty
PFACT (AVANTOR) make Micropipettes Cat No.- PFACAFSK-2	STARTER KIT- 5 FAB PIPETTES: Consists of 5 Micropipettes- 0.5-10 µl pipette 500-5000 µl pipette 10-100 µl pipette 20-200 µl pipette 100-1000 µl pipette 10, 200 & 1000 µl (Sterile tip-1 rack each)	01

Dr. Dipankar Chakraborti
Head of the Department
Department of Genetics,
University of Calcutta, Kolkata

Dr. Dipankar Chakraborti
Head of the Department
Department of Genetics
University of Calcutta



University of Calcutta
35, Ballygunge Circular Road
Kolkata – 700 019
Tel: +91-33-2461 5445

QUOTATION NOTICE

Ref. No.: Genet/BI(97)/LERC11/22-23

Dated: 01.08.2022

Quotations for the following item are invited from authorized parties for Department of Genetics. The cost to be incurred from Departmental BI grant. Suppliers are hereby requested to submit their quotation including the current GST registration certificate, PAN no., copy of registration certificate with statutory authorities and other credentials **within 5 working** days from the publication of this notice in a sealed envelope addressed to the undersigned. **Please provide a tender complete in all aspects, otherwise, it won't be considered.** The authority reserves the right to accept/reject any quotation without showing any reason thereof.

Item	Description	Qty
DRY BATH INCUBATOR	<ul style="list-style-type: none">Electrically operated to work on 220/230 volts, single phase, AC mains.Dry bath is made of Mild Steel sheet with powder coating paint.Single block is made of Aluminum with 36 nos. tube hole.Capacity: 1.5ml. x 18Nos. and 2.0ml. x 18 Nos.Temperature range above ambient to 150°C of an accuracy $\pm 1^{\circ}\text{C}$. and temperature controlled by SELECT Make Digital Temperature indicator cum Controller with PT-100 SensorComplete with Off/On Switch.	01

Dr. Dipankar Chakraborti
Head of the Department
Department of Genetics,
University of Calcutta, Kolkata

Dr. Dipankar Chakraborty
Associate Professor
Department of Genetics
University of Calcutta



University of Calcutta
35, Ballygunge Circular Road
Kolkata – 700 019
Tel: +91-33-2461 5445

QUOTATION NOTICE

Ref. No.: Genet/BI(97)/LERC12/22-23

Dated: 01.08.2022

Quotations for the following item are invited from authorized parties for Department of Genetics. The cost to be incurred from Departmental BI grant. Suppliers are hereby requested to submit their quotation including the current GST registration certificate, PAN no., copy of registration certificate with statutory authorities and other credentials **within 5 working days** from the publication of this notice in a sealed envelope addressed to the undersigned. **Please provide a tender complete in all aspects, otherwise, it won't be considered.** The authority reserves the right to accept/reject any quotation without showing any reason thereof.

No	Material Description.	Qty
1	Digital INCUBATOR WITH SHAKER Specification: Double walled with inner chamber of stainless steel & outer wall of M.S. Sheet duly Powder coated. Temperature will be controlled by Digital controller with air circulation. Temperature range : 5°C to maximum 60°C with accuracy range of $\pm 2^\circ\text{C}$. Chamber Size: 120 lit (4 cft) : Shaker Speed : 20-180 RPM Shaker Size : 13*13 inch Capacity : 500 ml *5 /100 ml*5 & 1 lit *2 Number of shelves: one (1 Nos. Stainless Steel adjustable trays) Microprocessor based PID controller Temperature range between which will be controlled by means of a Digital Temperature Controller in conjunction with an electronic relay.	01

Yours faithfully,

Dr. Dipankar Chakraborti
Department of Genetics,
University of Calcutta, Kolkata